

I0CZZ000.272.3.2021

Wrocław, 12.07.2021

INFORMATION FOR ECONOMIC CONTRACTORS No. 2

It concerns the public procurement procedure conducted in the Offer Inquiry mode, the subject of which is *The delivery of the five identical accelerometers with the necessary auxiliary devices. The order is carried out under the project „EPOS – European Plate Observing System (EPOS-PL+)”, TASK 5, Grant agreement No. POIR.04.02.00-00-C005/19-00.*

The Contracting Authority answers the Economic Contractors' questions:

Question No 1

In Appendix 3 to the Inquiry "Agreement template" in § 9 sec. 1 of the contract, the Contracting Authority used the provision "[...] contractual penalties of 2% of the value of the undelivered object of the contract [...]". Please indicate what value will the penalty be charged, gross or net?

Reply: Contractual penalties are calculated on the gross value of the contract.

Question No 2

In Appendix 4 to the Inquiry "Technical parameters of the subject of the order" in line no. 1. The Contracting Authority specified the required parameters of the accelerometer as follows: "Supply voltage 5 - 36 VDC". Does the Employer allow the delivery of the device with a supply voltage of 9 - 18 VDC?

Reply: Yes, the Contracting Authority allows the device with the above parameters.

Question Nr 3

In Appendix 4 to the Inquiry "Technical parameters of the subject of the order" in line no. 14. The Contracting Authority specified the required parameters of the digitizer as follows: "Power supply 5 - 36 V". Does the Contracting Authority allow the delivery of the device with a supply voltage of 9 - 18 V?

Reply: Yes, the Contracting Authority allows the device with the above parameters.

In addition, the Contracting Authority informs about the postponement of the deadline for submitting offers to July 16 this year. until 9:00 am.

Rector
Prof. dr hab. inż. Jarosław Bosy